

## FRAMELESS GLASS CURTAINS LTD

Dear Sir/ Madame,

This letter is to confirm that the following equipment has been calibrated against utilising a parasitic method so that traceability to a national standard is ensured. Further certification is available on request

## Date of Calibration 19th of June 2024

CEN Solutions Ltd – Equipment Reference	Customer Equipment Reference	Master Reading	Customer Reading	Error	Pass/Fai l	Next Calibration due
TES 1319 Digital Thermometer Serial No 041230107 Certificate No 0525259: Expiry: 01/8/24	CEM DT610-B Digital Thermometer Register No 1	20º/c	19º Celsius	-1°/c	Pass	18/06/2025
		86º/c	86 Celsius	+/-0 °/c	Pass	18/06/2025
		128º/c	128º Celsius	+/-0 °/c	Pass	18/06/2025

Yours Sincerely

For and Behalf of... **CENSolutions (SE) Ltd.** 

Michael Laney

M.J.Laney

## **CENSolutions Limited**